

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Sayamalai

Details About Robust Analysis

Lab No 12803

Detail About Monthly Summary

Month August

Detail about SDI

Year 2023

Constituent Group Chemistry III

Click on the analyte to view Graphical Data

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	181 mg/dL	0.60	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.6 mg/dL	-1.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	113 mg/dL	-0.17	0.22
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

*P. Srinivasan*

**MEDICAL OFFICER**  
**GOVT PRIMARY HEALTH CENTRE**  
 SAIGALPETA, 634 592

Links