



VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Latheri

[Details About Request Analysis](#)

Lab No 12041

[Detail About Monthly Summary](#)

Month August

[Detail about SDI](#)

Year 2023

Constituent Group Chemistry II

[All Analyser Results](#)

[Print](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl. No	Parameter	Method	Instrument	Target	Mean	SD	CV	SI	SI <sup>2</sup>	SI <sup>3</sup>
1	GLUCOSE II	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	2.90	186 mg/dL	1.16	0.42
2	CREATININE II	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.7 mg/dL	-0.50	0.01
3	CHOLESTROL II	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	126 mg/dL	2.42	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within ±1.01 to ±2.00	Good
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

---



---



---

Maintained by  
Software Solutions

**MEDICAL OFFICER**  
Govt PRIMARY HEALTH CENTRE  
LATHERI - 632 202, KATPADI TA  
VELLORE DISTRICT

Dr.SANKAR GANESH M.B.B.S  
Block Medical Officer  
Govt. Community Health centre  
Vaduganthangal, K.V.Kuppam



Directorate of Public  
Health and Preventive Medicine

### LEGAL IDENTITY CERTIFICATE

This is to certify that the following Primary Health Centre is functioning under the Department of Public Health and Preventive Medicine, Government of Tamil Nadu.

Name of the District	Name of the Health Unit District	Name of the Block	Name of the Primary Health Centre
Vellore	Vellore	K.V.Kuppam	Latheri PHC

This certificate is issued for submitting the Application for the "NABL Medical Entry LEVEL Testing M(EL)T Labs Program"

Block Medical Officer  
Govt. Community Health centre  
Vaduganthangal