

	<u>CA</u>	LIBRAT	TON CER	RTIFICA	ATE	
CERTIFICATE NO: SBS/CL	/23/12155					
Issue Date			MEDICAL DEVICES Page 80-09-2023			Page No:1 of
SRF No & Date			SRF/23/00493-0001 & 29-09-2023			
Receipt Date			29-09-2023			
Calibration Date			29-09-2023			
Calibration Due			28-09-2024			
Customer Name & Address			20-03-2024			
GOVERNMENT PRIMARY I	HEALTH CENT	RE.				
PERUNAZHI-623115,RAMAI	NATHAPURAM	DISTRICT				
			vice Under Calib	ration (DUC)		
escription : SE	MI AUTO ANALY		Make & Model		225 1990/70000000000000000000000000000000000	
National Control of the Control of t	MULTI		Sr. No	;	ALPHA TECHNOLOGIES & CHEM 100	
desolution : MU	LTI	1			BCAA17062655	
UC Condition : SA	ISFACTORY		dentification No	1	NA	
		nmental Condi	_ocation tions & Calibration	: D 1	LABORATATORY	
nvironmental Details	Temperature	25.6 ° C				
Calibration Procedure No SBS/CP/MD/20			Relative Humidity 52% RH Calibration done at ONSITE			
	1		nce Standards D		ONSITE	
.No Description Make/SI No:		Make/ SI No:	ince Standards D			
Floatrical Oct 1 A				Certificate N		Validity
		Rigel Medical & 44L-1059		M-230809-16-4		10-08-2024

RESULTS	3	ELECTRICAL SAFETY	
S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance >20MΩ	Measured values in M Ω	Uncertainty in % (±)
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in μA 209	13.92 Uncertainty in % (±)
3	Enclosure Leakage	Measured values in μA	7.3 Uncertainty in % (±)
25114 21/2	<500µAfor B,BF,CF	229	7.2

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards. Calibrated By

(Calibration Engineer) M.RAGUL



•	Technical Manager
	C.SHANMUGARAJ

Chief Executive

Authorised Signatory

SUNSHINE BIOMEDICAL SOLUTIONS