



CALIBRATION CERTIFICATE- SEMI AUTO ANALYSER

Hospital Name : Urban Primary Health Centre.
Address : Saidapet, Vellore.
Certificate No : SBS/CL/18-19/1526.

EQUIPMENT DETAILS			
Make	ROBONIK	Serial No	AT0681214RBK
Model	NA	Department	LAB
Calibration Date	03-07-2018	Next Due On	02-07-2019

CALIBRATION EQUIPMENT DETAILS	
Instrument Name	Electrical Safety Analyzer
Make	Fluke
Model	ESA 615
Serial No	2244202
Calibration Date	15-11-2017
Calibration Due	15-11-2018
Temperature	24 ° C
Humidity(40-70% RH)	58%

REMARKS : REPORT ATTACHED

ELECTRICAL SAFETY TEST PASSED AS PER IEC STANDARD 60601-1 FOR OUR MATER EQUIPMENT

CALIBRATION PASSED : YES NO

The above mentioned instrument has been calibrated using standards manufacturer recommended protocols, using equipments having traceability to National/International Standards. Devices for which there are no National/International Standards calibration standards are measured against in-house performance standard using accepted test procedures.
Partial reproduction of the certificate is not permitted

Calibrated By
(Mr. ANANDAPPAN)

Authorized Signatory
(Mr. SHAHMUGARAJ)

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First floor , Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087.