

CALIBRATION CERTIFICATE- CENTRIFUGE

Hospital Name : Tybav Primary Health Centre.
 Address : Saidapet, Vellore.
 Certificate No : SRS/CL/18-19/1525.

EQUIPMENT DETAILS			
Make	REMI	Serial No	CENTRIFUGE-01
Model	NA	Department	LAB
Calibration Date	03-07-2018	Next Due On	02-07-2019

CALIBRATION EQUIPMENT DETAILS		
Instrument Name	Electrical Safety Analyzer	Digital Tachometer
Make	Fluke	Systems
Model	ESA 615	HTM-590
Serial No	2244202	16-913
Calibration Date	15-11-2017	29-08-2017
Calibration Due	15-11-2018	29-08-2018
Temperature	24 ° C	-
Humidity(40-70% RH)	58%	-

REMARKS : REPORT ATTACHED

ELECTRICAL SAFETY TEST PASSED AS PER IEC STANDARD 60601-1 FOR OUR MATER EQUIPMENT

CALIBRATION PASSED : YES NO

The above mentioned instrument has been calibrated using standards manufacturer recommended protocols, using equipments having traceability to National/International Standards. Devices for which there are no National/International Standards calibration standards are measured against in-house performance standard using accepted test procedures.
 Partial reproduction of the certificate is not permitted

Calibrated By
 (Mr. SANDIPPAN)

Authorized Signatory
 (Mr. SHANMUGARAJ)

