

CALIBRATION DATA REPORT

| Electrical Safety Test           |                         |   |
|----------------------------------|-------------------------|---|
| Parameters Used                  | Measured in Test Gadget | Limits                                    |
| Voltage N-E                      | 2.6                     | < 3V                                      |
| Maximum Current Taken by the DUT | 0.1                     | As per Manufacture specification          |
| Earth Leakage Current NC         | 143                     | <500 $\mu$ A for B,BF,CF                  |
| Earth Leakage Current SFC        | 215                     | <1000 $\mu$ A for B,BF,CF                 |
| Enclosure Leakage Current NC     | 4.1                     | <100 $\mu$ A for B,BF,CF                  |
| Enclosure Leakage Current SFC    | 4.1                     | <500 $\mu$ A for B,BF,CF                  |
| Patient Leakage Current NC       | 4.2                     | <100 $\mu$ A for B,BF, <10 $\mu$ A for CF |
| Patient Leakage Current SFC      | 4.3                     | <500 $\mu$ A for B,BF, <50 $\mu$ A for CF |
| Insulation Resistance M-P-E      | Good                    | More than 100Mohm                         |
| Insulation Resistance M-A-P      | Good                    | More than 100Mohm                         |
| Insulation Resistance A-P-P-E    | Good                    | More than 100Mohm                         |

CALIBRATION RESULT:  CORRECTIVE MAINTENANCE REQUIRED  REMOVED FROM USE  ACCEPTABLE FOR USE

CALIBRATION ENGINEERS SIGN:

DATE: 03-07-2015