

Calibration Certificate,


Name		PRIMARY HEALTH CENTRE - MANSAHU	
Hospital			
Date of Installation/Calibration		01.03.2023	
Instrument Identification		Manshu's 100-1000 P.	
Serial no.	Date of Calibration	01.03.2023	CC Passed Yes <input checked="" type="checkbox"/>
Model	Calibration Due Date	30.07.2023	Due

Inspection LIST (Relevant information should be filled in)

	YES/NO
1. The Voltage of Instrument should be 220 V \pm 10%.	/
2. Ground Voltage should be less than 10 volts.	
3. The instrument is tested with the stimulator with all leads connected.	
4. AC frequency should be 50 Hz and all filters are in ON condition	
5. Venous pre-diluted modes are working OK	
6. All values working good	

Quality Control Result	Result	Remark
Checked instrument with Quality control, working Satisfactorily	OK.	
Test results is coming ok	✓	Remarks Done

Signature of Technical person



ERIZ MEDICARE PRIVATE LIMITED
 No. 19, SHANMUGA NAGAR,
 2nd MAIN ROAD, MANNIVAYAL,
 CHENNAI - 600 044
 CELL: 95008 46575

