

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name

Govt. PHC, Nerumbur

Lab No

12251

Constituent Group

Chemistry III

Date of Result Entered :

17/03/2023

PT item

Lyophilized human serum based

Date of Report Published:

04/04/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of	DV	Participants Your SEA				
		Name	,	Participants		CA	SD	Value		
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	87 mg/dL	-0.66	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.4 mg/dL	0.58	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	114 mg/dL	0.53	0.28
	SDI F	Range		Inte	rpretatio	n			2251	

SDI Range	Interpretation	12251
Within -1.00 to +1.00	Excellent.	, 17/03/
Within ±1.01 to ±2.00	Good.	
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.	illaria indo-
Beyond ±3.0	Unacceptable performance. Action Signal.	

LAB ADDRESS:

Govt. PHC, Nerumbur THE MEDICAL OFFICER. Chengalpattu

TAMILNADU

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102 Panela Christudos

Dr. Pamela Christudoss **CMC EQAS Coordinator**

mg/dL

Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed. Data in CMC EQAS reports is confidential CMC EQAS does not sub contract any components ******* End of Report ******

Vellora