



**CHRISTIAN MEDICAL COLLEGE**  
DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name: Govt. PHC, Kombanaiipudur      Lab No: 12211  
 Constituent Group: Chemistry III      Date of Result Entered: 19/04/2023  
 PT Item: Lyophilized human serum based      Date of Report Published: 06/05/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	121 mg/dL	0.21	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2 mg/dL	-0.91	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	102 mg/dL	-1.66	0.28
<b>SDI Range</b>										
Within -1.00 to +1.00										
Within ±1.01 to ±2.00										
Within ±2.01 to ±2.99										
Beyond ±3.0										
<b>Interpretation</b>										
Excellent.										
Good.										
Accept with caution. Warning Signal.										
Unacceptable performance. Action Signal.										

**LAB ADDRESS :**  
Govt. PHC, Kombanaiipudur  
THE MEDICAL OFFICER, KODUMUDI  
Kombanaiipudur  
TAMILNADU638152

**Coordinator Contact Details:**  
Email: cilingc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*[Signature]*  
**MEDICAL OFFICER,**  
GOVT PRIMARY HEALTH CENTRE,  
KOMBANAI PUDUR, UNJALUR VIA,  
TAMILNADU-638152



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name Govt. PHC, Kombanaipudur Lab No 12211  
Constituent Group Chemistry III Date of Result Entered : 17/06/2023  
PT Item Lyophilized human serum based Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	285.9 mg/dL	-0.05	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.6 mg/dL	0.18	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	101.6 mg/dL	-2.39	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

#### LAB ADDRESS :

Govt. PHC, Kombanaipudur  
THE MEDICAL OFFICER, KODUMUDI  
Kombanaipudur  
TAMILNADU638152

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*P.P.*  
அரசாங்கத்தவர்  
அரசு ஆரம்பகல்பகல்ப நல்லைம்  
சென்மல்பகல்பகல்ப  
ஈரோடுமல்பகல்ப 638152.



**CHRISTIAN MEDICAL COLLEGE**  
 DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - JULY 2023



Lab Name: Govt. PHC, Kombanaiipudur  
 Constituent Group: Chemistry III  
 PT Item: Lyophilized human serum based  
 Lab No: 12211  
 Date of Result Entered: 20/07/2023  
 Date of Report Published: 07/08/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	106 mg/dL	-0.55	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	6 mg/dL	0.49	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	104 mg/dL	-1.93	0.22
SDI Range Interpretation										
Within -1.00 to +1.00 Excellent.										
Within ±1.01 to ±2.00 Good.										
Within ±2.01 to ±2.99 Accept with caution. Warning Signal.										
Beyond ±3.0 Unacceptable performance. Action Signal.										

LAB ADDRESS :  
 Govt. PHC, Kombanaiipudur  
 MEDICAL OFFICER, KODUMUDI  
 Kombanaiipudur  
 TAMILNADU638152

Coordinator Contact Details:  
 Email: clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Kombanalpudur

Details About Robust Analysis

Lab No 12211

Detail About Monthly Summary

Month August

Detail about SDI

Year 2023

Constituent Group Chemistry III

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	163 mg/dL	-1.40	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.6 mg/dL	-1.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	107 mg/dL	-1.47	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



**CHRISTIAN MEDICAL COLLEGE**  
 DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - JULY 2023



Lab Name: Govt. PHC, Kombanaiapur  
 Lab No: 12211  
 Constituent Group: Chemistry III  
 Date of Result Entered: 20/07/2023  
 PT Item: Lyophilized human serum based  
 Date of Report Published: 07/08/2023

S.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	106 mg/dL	-0.55	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	6 mg/dL	0.49	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	104 mg/dL	-1.93	0.22
SDI Range Interpretation										
Within -1.00 to +1.00 Excellent.										
Within ±1.01 to ±2.00 Good.										
Within ±2.01 to ±2.99 Accept with caution. Warning Signal.										
Beyond ±3.0 Unacceptable performance. Action Signal.										

LAB ADDRESS :  
 Govt. PHC, Kombanaiapur  
 DISTRICT MEDICAL OFFICER, KODUMUDI  
 Kombanaiapur  
 TAMILNADU638152

*Pamela Christudoss*

Coordinator Contact Details:  
 Email:clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*

VIEW LAB YEARLY SUMMARY

**CHRISTIAN MEDICAL COLLEGE**  
**CMC EXTERNAL QUALITY ASSURANCE SCHEME**  
**2022 EQAS YEARLY SUMMARY**

The yearly summary is based on the overall mean SDI of every analyte. The groups are made depending upon the number of results received for every individual analyte. The rank is mentioned for every analyte with in the group. This modification in the yearly summary is done with the guidelines provided by NABL during the assessment for accreditation.

Group I : Laboratories who reported 10 to 12 months during 2022

Group II : Laboratories who reported 8 to 9 months during 2022

Group III : Laboratories who reported less than 6 months during 2022

Sl. No.	Parameter	Rank Group	No. of Labs	OMSDI	Grade
1	GLUCOSE III	Group I	1941	0.94	A
2	CREATININE III	Group I	1927	0.90	A
3	CHOLESTROL III	Group I	1923	1.07	B

Only first 2 Groups are Graded


OMSDI Range	Grade	Remarks
0.00 to 1.00	A	Excellent
1.01 to 2.00	B	Good
2.01 to 2.99	C	Satisfactory
Beyond ±3.00	D	Unacceptable

Your Lab No 12211

Constituent Group Chemistry III

Address

Govt. PHC, Kombanaiapur,  
 THE MEDICAL OFFICER, KOCUMUDI,  
 Kombanaiapur,  
 TAMILNADU, 638152.

  
 Dr. S. Srinivasan  
 Dr. S. Srinivasan  
 Dr. S. Srinivasan  
 Dr. S. Srinivasan 638152

**Links**

- Modify Address
- Modify Methods
- Submit Lab Results
- View Lab Monthly Summary
- Yearly Summary
- Change Password
- Home
- Logout

Print

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Kombanalpudur  
 Lab No 12211  
 Month May  
 Year 2023  
 Constituent Group Chemistry III

Details About Robust Analysis  
 Detail About Monthly Summary  
 Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 18/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	178.2 mg/dL	-0.29	0.44
2	CREATININE III	Jaffes Klnetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	108.9 mg/dL	-0.98	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

அரசு ஆராய்ச்சித் துறை  
 அரசு ஆராய்ச்சித் துறை நிலையம்  
 கொம்பலைப்புதுர  
 ஈரோடு மாவட்டம் 638152.