

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Manimuthar	Details About Robust Analysis
Lab No	12765	Detail About Monthly Summary
Month	August	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	175 mg/dL	-0.06	0.42
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	174	28.29	15.20	4.30	27 mg/dL	-0.30	0.65
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	107 mg/dL	-1.47	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

MEDICAL OFFICER
 GOVT. PHC, MANIMUTHAR
 ARUNACHAL PRADESH
 INDIA-786001