




ROOTS INDUSTRIES INDIA LIMITED
ROOTS METROLOGY & TESTING LABORATORY
 R.K.G Industrial Estate, Ganapathy, Coimbatore-641 006,India
 Phone :9003432356



CALIBRATION CERTIFICATE

ULR Number: CC275219000083790F

Name and address of the customer The Medical Officer Government Primery Health Centre melattur–thanjavur (dist)	Certificate no : Rmtl/01/319107852/22
	Date of calibration :1/10/2023
	Suggested calibration due : 30/09/2024
	Customer po .reference.& date:

1.DESRIPTION OF DUC

Name Of The Duc: pipette Make : THERMO SCIENTIFIC Model : FINEPIPETTE Sl.No : rw13159 Range/Size : 10-100 micro litter	Receiving Condition Of Duc :Good Ref. Document :As Per Manufactures Specification Environmental Condition :25.3 Deg & 54.3% Rh Uncertainty Of Measurement :As Per Report Location :lab
---	--

THE MASTERS ARE TRACEABLE TO NATIONAL/INTERNATIONAL STANDARDS AS PER ISO/IEC 17025.

2.DETAILES OF MASTER USED

Standard Instrument Used	Make/ Model	Id No	Traceable To/Certificate No	Validity
Multi Product Calibrator	Fluke	Rml-262	Wavecal,Blr/Cs-18/01-01	1-6-2024
Micro balance	mettler toledo/XP26PC	QP-01	B115129039/CC-2523	1-6-2024
Universal Measuring Instrument With Sensor For Envirmental Parameter	ALMEMO/2590-4AS	QP-10	H19070544/CC-2052	1-6-2024

CALIBRATION FIELD: WEIGHT Calibration

Sl.no	Range	Master readings micro litter	Calibrated values micro litter	Error observed micro litter	Uncertainty
1	10-100 micro litter	10	9.9	0.1	0.055
2	10-100 micro litter	50	49.9	0.1	0.055
3	10-100 micro litter	80	79.9	0.1	0.055
4	10-100 micro litter	90	89.9	0.1	0.055
5	10-100 micro litter	100	99.9	0.1	0.055

4.remarks

- 4.1 uncertainty estimated for a confidence level of 95% at the k factor=2
 - 4.2the reported results are valid only for the condition of the inst./gauge received at the time of and under the condition of the calibration
 - 4.3 DUC –device under calibration
- Calibration label /status : Calibration result pass & fail are indicated

 CALIBRATION ENGINEER	 CALIBRATION SEAL	 AUTHORIZED SIGN	This Report Refers Only To The Sample Submitted And May Not Be Reproduces In Full Or In Part, Without Written Permission From Roots Metrology & Testing Laboratory, Coimbatore,
--------------------------	----------------------	---------------------	---



ROOTS INDUSTRIES INDIA LIMITED
ROOTS METROLOGY & TESTING LABORATORY
 R.K.G Industrial Estate, Ganapathy, Coimbatore-641 006,India
 Phone :9003432356



NABL/ILAC/024



NABL ACCREDITATION
 CERTIFICATE NO CC-2752

CALIBRATION CERTIFICATE

ULR Number: CC275219000083790F

Name and address of the customer The Medical Officer Government Primery Health Centre melattur–thanjavur (dist)	Certificate no : Rmtl/01/319107852/23 Date of calibration :1/10/2023 Suggested calibration due : 30/09/2024 Customer po .reference.& date:
---	---

1.DESRIPTION OF DUC

Name Of The Duc: pipette Make : THERMO SCIENTIFIC Model : FINEPIPETTE Sl.No : NA Range/Size : 100-1000 micro litter	Receiving Condition Of Duc :Good Ref. Document :As Per Manufactures Specification Environmental Condition :25.3 Deg & 54.3% Rh Uncertainty Of Measurement :As Per Report Location :lab
---	--

THE MASTERS ARE TRACEABLE TO NATIONAL/INTERNATIONAL STANDARDS AS PER ISO/IEC 17025.

2.DETAILES OF MASTER USED

Standard Instrument Used	Make/ Model	Id No	Traceable To/Certificate No	Validity
Multi Product Calibrator	Fluke	Rml-262	Wavecal,Blr/Cs-18/01-01	1-6-2024
Micro balance	mettle toledo/XP26PC	QP-01	B115129039/CC-2523	1-6-2024
Universal Measuring Instrument With Sensor For Envirmental Parameter	ALMEMO/2590-4AS	QP-10	H19070544/CC-2052	1-6-2024

CALIBRATION FIELD: WEIGHT Calibration

Sl.no	Range	Master readings micro litter	Calibrated values micro litter	Error observed micro litter	Uncertainty
1	100-1000 micro litter	100	99.9	0.1	0.055
2	100-1000 micro litter	500	499.9	0.1	0.055
3	100-1000 micro litter	800	799.9	0.1	0.055
4	100-1000 micro litter	900	899.9	0.1	0.055
5	100-1000 micro litter	1000	999.9	0.1	0.055

4.remarks

4.1 uncertainty estimated for a confidence level of 95% at the k factor=2

4.2the reported results are valid only for the condition of the inst./gauge received at the time of and under the condition of the calibration

4.3 DUC –device under calibration

Calibration label /status : Calibration result pass & fail are indicated

 CALIBRATION ENGINEER	 CALIBRATION SEAL	 AUTHORIZED SIGN	This Report Refers Only To The Sample Submitted And May Not Be Reproduces In Full Or In Part, Without Written Permission From Roots Metrology & Testing Laboratory, Coimbatore,
--------------------------	----------------------	---------------------	--



ROOTS INDUSTRIES INDIA LIMITED
ROOTS METROLOGY & TESTING LABORATORY
 R.K.G Industrial Estate, Ganapathy, Coimbatore-641 006,India
 Phone :9003432356



NABL/ILAC/024

NABL ACCREDITATION
 CERTIFICATE NO CC-2752

CALIBRATION CERTIFICATE

ULR Number: CC275219000083790F

Name and address of the customer The Medical Officer Government Primary Health Centre melattur–thanjavur (dist)	Certificate no : Rmtl/01/319107852/24 Date of calibration :1/10/2023 Suggested calibration due : 30/09/2024 Customer po .reference.& date:
---	---

1.DESRIPTION OF DUC

Name Of The Duc: CENTRIFUGE Make :lincosci inst Model : mk-10 Sl.No : 138/385 Range/Size : 1-4000RPM/230v	Receiving Condition Of Duc :Good Ref. Document :As Per Manufactures Specification Environmental Condition :25.3 Deg & 54.3% Rh Uncertainty Of Measurement :As Per Report Location :lab
--	---

THE MASTERS ARE TRACEABLE TO NATIONAL/INTERNATIONAL STANDARDS AS PER ISO/IEC 17025.

2.DETAILED OF MASTER USED

Standard Instrument Used	Make/ Model	Id No	Traceable To/Certificate No	Validity
Decade Megohm	Vaisashika	Rml-230	Vaiseshika,Ambala Cantt/Ved/Nabl/18-19(Et)/92	1-6-2024
Digital Tachometer	Mextec	Rml-251	Gsg/19/206/Tm/01	1-6-2024
Multi Product Calibrator	Fluke	Rml-262	Wavecal,Blr/Cs-18/01-01	1-6-2024
Multi Product Calibrator (Ac/Dc Current Coil)	Transmille	Rml-252-1	Wavecal,Blr/Cs-18/69-01	1-6-2024
Leakage Current Tester	Htc/CI-2054	Rml-254-1	Rmtl/01/319107848	1-6-2024

CALIBRATION FIELD: Electro Technical (Voltage Calibration -Ac)

Sl.no	Range	Master readings v	Calibrated values v	Error observed v	Uncertainty v
1	0 v to 400v	230	231	1	0.055

CURRENT CALIBRATIONAc Current @50Hz

Sl.no	Range	Master readings mA	Calibrated values mA	Error observed mA	Uncertainty mA
1	40Ma-100A	2	2.002	-0.002	0.045

Resistance calibration

Sl.no	Range	Master readings ohms	Calibrated values ohms	Error observed ohms	Uncertainty Ohms
1	40	39	39.01	0.001	0.007

CALIBRATION FIELD: RPM Calibration

Sl.no	Range	Master readings RPM	Calibrated values RPM	Error observed RPM	Uncertainty RPM
1	1RPM to 4000RPM	1600RPM	1599RPM	1RPM	0.055

4.remarks

4.1 uncertainty estimated for a confidence level of 95% at the k factor=2

4.2the reported results are valid only for the condition of the inst./gauge received at the time of and under the condition of the calibration

4.3 DUC –device under calibration

Calibration label /status : Calibration result pass & fail are indicated

 CALIBRATION ENGINEER	 CALIBRATION SEAL	 AUTHORIZED SIGN	This Report Refers Only To The Sample Submitted And May Not Be Reproduces In Full Or In Part, Without Written Permission From Roots Metrology & Testing Laboratory, Coimbatore,
--------------------------	----------------------	---------------------	---



ROOTS INDUSTRIES INDIA LIMITED
ROOTS METROLOGY & TESTING LABORATORY
 R.K.G Industrial Estate, Ganapathy, Coimbatore-641 006,India
 Phone :9003432356



CALIBRATION CERTIFICATE

ULR Number: CC275219000083790F

Name and address of the customer The Medical Officer Government Primary Health Centre melattur–thanjavur (dist)	Certificate no : Rmtl/01/319107852/25
	Date of calibration :1/10/2023
	Suggested calibration due : 30/09/2024
	Customer po .reference.& date:

1.DESRIPTION OF DUC

Name Of The Duc: semi auto analyser Make : ROBONICK Model : PRIETEST TOUCH Sl.No : ATCD3891220RBK Range/Size : 230v	Receiving Condition Of Duc :Good Ref. Document :As Per Manufactures Specification Environmental Condition :25.3 Deg & 54.3% Rh Uncertainty Of Measurement :As Per Report Location :lab
--	---

THE MASTERS ARE TRACEABLE TO NATIONAL/INTERNATIONAL STANDARDS AS PER ISO/IEC 17025.

2.DETAILED OF MASTER USED

Standard Instrument Used	Make/ Model	Id No	Traceable To/Certificate No	Validity
Decade Megohm	Vaisashika	Rml-230	Vaiseshika,Ambala Cantt/Ved/Nabl/18-19(Et)/92	1-6-2024
Multi Product Calibrator	Fluke	Rml-262	Wavecal,Blr/Cs-18/01-01	1-6-2024
Multi Product Calibrator (Ac/Dc Current Coil)	Transmille	Rml-252-1	Wavecal,Blr/Cs-18/69-01	1-6-2024
Leakage Current Tester	Htc/CI-2054	Rml-254-1	Rmtl/01/319107848	1-6-2024

CALIBRATION FIELD: Electro Technical (Voltage Calibration -Ac)

Sl.no	Range	Master readings v	Calibrated values v	Error observed v	Uncertainty v
1	0 v to 400v	230	231	1	0.055

CURRENT CALIBRATIONAc Current @50Hz

Sl.no	Range	Master readings mA	Calibrated values mA	Error observed mA	Uncertainty mA
1	40Ma-100A	2	2.002	-0.002	0.045

Resistance calibration

Sl.no	Range	Master readings ohms	Calibrated values ohms	Error observed ohms	Uncertainty Ohms
1	40	39	39.01	0.001	0.007

4.remarks

- 4.1 uncertainty estimated for a confidence level of 95% at the k factor=2
 - 4.2the reported results are valid only for the condition of the inst./gauge received at the time of and under the condition of the calibration
 - 4.3 DUC –device under calibration
- Calibration label /status : Calibration result pass & fail are indicated

 CALIBRATION ENGINEER	 CALIBRATION SEAL	 AUTHORIZED SIGN	This Report Refers Only To The Sample Submitted And May Not Be Reproduces In Full Or In Part, Without Written Permission From Roots Metrology & Testing Laboratory, Coimbatore,
--------------------------	----------------------	---------------------	---



ROOTS INDUSTRIES INDIA LIMITED
ROOTS METROLOGY & TESTING LABORATORY
 R.K.G Industrial Estate, Ganapathy, Coimbatore-641 006,India
 Phone :9003432356



NABL/ILAC/024

NABL ACCREDITATION
 CERTIFICATE NO CC-2752

CALIBRATION CERTIFICATE

ULR Number: CC275219000083790F

Name and address of the customer The Medical Officer Government Primary Health Centre melattur–thanjavur (dist)	Certificate no : Rmtl/01/319107852/26 Date of calibration : 1/10/2023 Suggested calibration due : 30/09/2024 Customer po .reference.& date:
---	--

1.DESRIPTION OF DUC

Name Of The Duc: MICROSCOPE Make : LABOMED Model : BD 13 Sl.No : 9437185 Range/Size : 230v	Receiving Condition Of Duc : Good Ref. Document : As Per Manufactures Specification Environmental Condition : 25.3 Deg & 54.3% Rh Uncertainty Of Measurement : As Per Report Location : lab
---	--

THE MASTERS ARE TRACEABLE TO NATIONAL/INTERNATIONAL STANDARDS AS PER ISO/IEC 17025.

2.DETAILED OF MASTER USED

Standard Instrument Used	Make/ Model	Id No	Traceable To/Certificate No	Validity
Decade Megohm	Vaisashika	Rml-230	Vaiseshika,Ambala Cantt/Ved/Nabl/18-19(Et)/92	1-6-2024
Multi Product Calibrator	Fluke	Rml-262	Wavecal,Blr/Cs-18/01-01	1-6-2024
Multi Product Calibrator (Ac/Dc Current Coil)	Transmille	Rml-252-1	Wavecal,Blr/Cs-18/69-01	1-6-2024
Leakage Current Tester	Htc/CI-2054	Rml-254-1	Rmtl/01/319107848	1-6-2024

CALIBRATION FIELD: Electro Technical (Voltage Calibration -Ac)

Sl.no	Range	Master readings v	Calibrated values v	Error observed v	Uncertainty v
1	0 v to 400v	230	231	1	0.055

CURRENT CALIBRATION Ac Current @50Hz

Sl.no	Range	Master readings mA	Calibrated values mA	Error observed mA	Uncertainty mA
1	40Ma-100A	2	2.002	-0.002	0.045

Resistance calibration

Sl.no	Range	Master readings ohms	Calibrated values ohms	Error observed ohms	Uncertainty Ohms
1	40	39	39.01	0.001	0.007

4.remarks

- 4.1 uncertainty estimated for a confidence level of 95% at the k factor=2
 - 4.2the reported results are valid only for the condition of the inst./gauge received at the time of and under the condition of the calibration
 - 4.3 DUC –device under calibration
- Calibration label /status : Calibration result pass & fail are indicated

 CALIBRATION ENGINEER	 CALIBRATION SEAL	 AUTHORIZED SIGN	This Report Refers Only To The Sample Submitted And May Not Be Reproduces In Full Or In Part, Without Written Permission From Roots Metrology & Testing Laboratory, Coimbatore,
--------------------------	----------------------	---------------------	---