



CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/12171

Page. No : 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE, MELAKIDARAM-623528, RAMNAD DISTRICT.	SRF No.	SRF/23/00438-0002
	SRF Date	30-09-2023
	Date of Receipt	29-09-2023
	Date of Calibration	30-09-2023
	Due Date for Calibration	29-09-2024
	Issue Date	30-09-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	THERMO SCIENTIFIC
Range	100-1000µl	Model	FINNPIPETTE F3
Resolution	1µl	Material	PVC
Serial Number	QW14074	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.8	Humidity (%RH)	53	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

Sl. No.	Description	ID.No. / Sl. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl										
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	100.51	100.52	100.53	100.52	100.51	100.50	0.50	0.02	0.47
		100.49	100.48	100.49	100.49	100.48				
2	500	500.40	500.42	500.43	500.44	500.46	500.43	0.43	0.02	0.47
		500.45	500.44	500.43	500.42	500.41				
3	1000	1001.24	1001.26	1001.23	1001.22	1001.22	1001.22	1.22	0.03	0.47
		1001.18	1001.17	1001.20	1001.24	1001.26				

Remarks

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
- 6.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

(Calibration Engineer)
 MADHIBAN



Authorised by:

(Quality Manager/Chief Executive)
 C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/12172

Page. No: 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE, MELAKIDARAM-623528,RAMNAD DISTRICT.	SRF No.	SRF/23/00438-0003
	SRF Date	30-09-2023
	Date of Receipt	29-09-2023
	Date of Calibration	30-09-2023
	Due Date for Calibration	29-09-2024
	Issue Date	30-09-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	THERMO SCIENTIFIC
Range	10-100µl	Model	FINNPIPETTE F3
Resolution	0.2µl	Material	PVC
Serial Number	QW13600	Operating Range	10-100µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.8	Humidity (%RH)	53	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl											
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)	
1	10.0	9.87	9.86	9.87	9.86	9.87	9.86	-0.14	0.01	0.47	
		9.86	9.82	9.86	9.87	9.86					
2	50.0	49.46	49.47	49.48	49.47	49.46	49.48	-0.52	0.01	0.47	
		49.47	49.48	49.48	49.49	49.49					
3	100.0	99.37	99.38	99.39	99.31	99.40	99.40	-0.60	0.04	0.47	
		99.42	99.42	99.44	99.44	99.45					

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

M. Adhiban
(Calibration Engineer)
M.ADHIBAN



Authorised by:

C. Sivabalan
(Quality Manager/Chief Executive)
C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.