



CALIBRATION REPORT	
Report No: HBMS/CAL/23-24	Calibration Date: 15.09.2023
	Calibration Valid: 14.03.2024
<b>1.1. CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, GOVT.UPGRADED PRIMARY HEALTH CENTRE, NALLUR, PARAMATHI BLOCK, NAMAKKAL-637203.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	PIPETTE
I Manufactured by	THERMOSCIENTIFIC
II Model	FINNIPIPETTE F3
III Value	100µl
IV Serial No	GUPHC/NLR/PIP/003

**Calibration standards used :**


Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)
	100µl
1	100.05
2	100.06
3	100.01
4	99.97
5	100.01
6	100.02

For Heart Beat Medical Systems,

(Authorized signatory)




MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE - 04, hbmedicalsystem@gmail.com