



ROOTS INDUSTRIES INDIA LIMITED
ROOTS METROLOGY & TESTING LABORATORY
 R.K.G Industrial Estate, Ganapathy, Coimbatore-641 006,India
 Phone :9003432356



NABL/ILAC/024

NABL ACCREDITATION
 CERTIFICATE NO CC-2752

CALIBRATION CERTIFICATE

ULR Number: CC275219000083790F

Name and address of the customer The Medical Officer Government Primary Health Centre vadaseri–thanjavur (dist)	Certificate no : Rmtl/01/319107852/21 Date of calibration :1/10/2023 Suggested calibration due : 30/09/2024 Customer po .reference.& date:
---	---

1.DESRIPTION OF DUC

Name Of The Duc: MICROSCOPE Make : LABOMED Model : BD 13 Sl.No : 21869 Range/Size : 230v	Receiving Condition Of Duc :Good Ref. Document :As Per Manufactures Specification Environmental Condition :25.3 Deg & 54.3% Rh Uncertainty Of Measurement :As Per Report Location :lab
---	---

THE MASTERS ARE TRACEABLE TO NATIONAL/INTERNATIONAL STANDARDS AS PER ISO/IEC 17025.

2.DETAILED OF MASTER USED

Standard Instrument Used	Make/ Model	Id No	Traceable To/Certificate No	Validity
Decade Megohm	Vaisashika	Rml-230	Vaiseshika,Ambala Cantt/Ved/Nabl/18-19(Et)/92	1-6-2024
Multi Product Calibrator	Fluke	Rml-262	Wavecal,Blr/Cs-18/01-01	1-6-2024
Multi Product Calibrator (Ac/Dc Current Coil)	Transmille	Rml-252-1	Wavecal,Blr/Cs-18/69-01	1-6-2024
Leakage Current Tester	Htc/CI-2054	Rml-254-1	Rmtl/01/319107848	1-6-2024

CALIBRATION FIELD: Electro Technical (Voltage Calibration -Ac)

Sl.no	Range	Master readings v	Calibrated values v	Error observed v	Uncertainty v
1	0 v to 400v	230	231	1	0.055

CURRENT CALIBRATIONAc Current @50Hz

Sl.no	Range	Master readings mA	Calibrated values mA	Error observed mA	Uncertainty mA
1	40Ma-100A	2	2.002	-0.002	0.045

Resistance calibration

Sl.no	Range	Master readings ohms	Calibrated values ohms	Error observed ohms	Uncertainty Ohms
1	40	39	39.01	0.001	0.007

4.remarks

- 4.1 uncertainty estimated for a confidence level of 95% at the k factor=2
 - 4.2the reported results are valid only for the condition of the inst./gauge received at the time of and under the condition of the calibration
 - 4.3 DUC –device under calibration
- Calibration label /status : Calibration result pass & fail are indicated

 CALIBRATION ENGINEER	 CALIBRATION SEAL	 AUTHORIZED SIGN	This Report Refers Only To The Sample Submitted And May Not Be Reproduces In Full Or In Part, Without Written Permission From Roots Metrology & Testing Laboratory, Coimbatore,
--------------------------	----------------------	---------------------	---