

CALIBRATION / TESTING REPORT							
CUSTOMER DETAILS : GOVT.PRIMARY HEALTH CENTRE CHOKKAMPATTI,TENKASI, TAMIL NADU-627751		CERTIFICATE NO TICS/2023/CHOK/TSI/03		EQUIPMENT NAME MICROPIPETTE			
		MANUFACTURE MODEL		THERMOFISHER SCIENTIFIC 10-100µL			
		SERIAL NO RW09433		INSTALLED DEPARTMENT		LAB	
CALIBRATION DATE 11/09/2023		CALIBRATION DUE DATE 10/09/2024					
1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No							
TEST EQUIPMENT		ID. NO		CERTIFICATE NUMBER		VALIDITY	
ELECTRONIC SEMI MICRO BALANCE		NCS/WB01		TSC/23-24/5424-1		27-JUNE-2024	
TEST REPORT						Temperature 25°C Humidity RH 55%	
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS		
VOLUME	µL	50	49.46	± 1	PASS		
		100	999.80	± 1	PASS		
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.							
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE							
Calibrated By : <i>Per</i>						Authorized Sign With Seal 	