



G MULTI SERVICES

78, Yahoob Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin: 638 452.
 gmultiservicesgobi@gmail.com. Cell : 96597 20654.
 Bio Medical Instruments Services

Er. A. Ibathul Raziq DECE., BE (ICE)
 Bio Medical Engineer

Calibration Protocol & Report for Medical Equipment

Equipment : PIPETTE
 MAKE : FINNE PIPETTE F3
 MODEL : 100-1000 μ L

PROTECT CLASS / TYPE : Risk Class II
 S. No. : VA/PE/001
 Location : UPHC, VANIAMBADI.

S.No	Measurement Parameters	Status	Remarks
1	Electrical Test (a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	----	-
2	Physical Condition Test (a) Cabinet & Visible damages, if any (b) Risk of mechanical damages, if any	No No	Checked OK
3	Performance Test (e) Mean 99.6 μ l (f) A -0.5 μ l (g) Acc % -0.71 (h) CV% -0.36	OK OK OK Ok	
4	Calibration Test min volume 100 μ l max Volume 1000 μ l (e) A+or - 1.5, 5 μ l (f) S 1.5 μ l max Vol. 1000 μ l (g) Acc+ or - 1.50, 0.50 (h) CV% 0.60, 0.20	Ok Ok Ok Ok	Tested
Formulated By: Edition Research Laboratory		Issued By:	
Endorsed By : Research team head		User Department:	
Reviewed On: 18.08.23 Next Review : 17.08.24			