

CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : GOVT.PRIMARY HEALTH CENTRE CHOKKAMPATTI,TENKASI, TAMIL NADU-627751		CERTIFICATE NO TICS/2023/CHOK/TSI/02		EQUIPMENT NAME MICROPIPETTE	
		MANUFACTURE MODEL 100-1000µL		THERMOFISHER SCIENTIFIC	
		SERIAL NO RW11956		INSTALLED DEPARTMENT LAB	
CALIBRATION DATE	11/09/2023				
CALIBRATION DUE DATE	10/09/2024				
1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					
TEST EQUIPMENT	ID. NO	CERTIFICATE NUMBER	VALIDITY		
ELECTRONIC SEMI MICRO BALANCE	NCS/WB01	TSC/23-24/5424-1	27-JUNE-2024		
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	100	100.1	± 1	PASS
		1000	999.3	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By : <i>Per</i>					
Authorized Sign With Seal 					