



CALIBRATION / TESTING REPORT

CUSTOMER DETAILS : GOVT.PRIMARY HEALTH CENTRE CHOKKAMPATTI,TENKASI, TAMIL NADU-627751		CERTIFICATE NO	TICS/2023/CHOK/TSI/01
		EQUIPMENT NAME	CENTRIFUGE
		MANUFACTURE	REMI
		MODEL	C 852
CALIBRATION DATE	11/09/2023	SERIAL NO	GALC 330
CALIBRATION DUE DATE	10/09/2024	INSTALLED DEPARTMENT	LAB

- 1. General maintenance required Yes No
- 2. Any Physical damage observed Yes No
- 3. Spare Replacement (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER
DIGITAL TACHOMETER	KUSAM MECO	KM2234 BL	S1049875
ELECTRICAL SAFTY ANALYZER	DATREND	V Pad-mini	VPM 18040011

TEST REPORT

Temperature 25°C
Humidity RH 55%

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
ROTATION PER MINUTE	RPM	Speed 4	2635	± 5	PASS
		Speed 5	3008		

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE

Calibrated By :



*Electrical safety test result attached