



# G MULTI SERVICES


78, Yahoob Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin: 638 452.  
gmultiservicesgobi@gmail.com. Cell : 96597 20654.  
Bio Medical Instruments Services

Er. A. Ibathul Raziq DECE., BE (ICE)  
Bio Medical Engineer

## Calibration Protocol & Report for Medical Equipment

Equipment : MICROSCOPE  
MAKE : OLYPUSCH201  
MODEL : NA

PROTECT CLASS /TYPE : Risk Class II/CF  
S. No. : I11LO22  
Location : UPHC, VANİYAMBADI.

S.No	Measurement Parameters	Status	Remarks
1	Electrical Test (a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	235 volts 0 to 5volts No Zero ohms 10mA ---	Stabilizer/Ups  With in Ranges Well
2	Physical Condition Test (a) Cabinet & Visible damages, if any (b) Power Supply unit, halogen lamps, Power cord accessories (c) Risk of mechanical damages, if any	No OK  No	  Checked ok
3	Performance Test (a) Halogen lamps intensity (b) Intensity test taken by lux meter for a particular lamp & results (c) Measurement of lux/Lens range (d) Lamp life time	Normal OK  OK OK	Fine Brightness Perfect 20000 hrs Apprx.
4	Calibration Test (a) LED lamps intensity (b) Intensity test taken by lux meter for a particular lamp & results (c) Measurement of Lux/Lenses cal (d) Lamp life time	OK Done  1200 Lux OK	In range Peak  in range In range
Formulated By: Edition Research Laboratory		Issued By:	
Endorsed By : Research team head			
Reviewed On: 18.08.23 Next Review : 17.08.24			