



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name Govt. PHC, Panruti Lab No 10535
Constituent Group Chemistry III Date of Result Entered : 17/05/2023
PT item Lyophilized human serum based Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	186 mg/dL	0.55	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.9 mg/dL	1.06	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	108 mg/dL	-1.13	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Panruti
CLINICAL LABORATORY, KANCHIPURAM(DIST)
Panruti
TAMILNADU631604

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

Dr. Pamela Christudoss
மருத்துவ அலுவலர்
அரசு ஆரம்ப சுகாதார நிலையம்
பனாருட்டி - 631604.
பரிசீலனா செய்து ஒன்றியம்
காஞ்சிபுரம் மாவட்டம்