

2023 - AUGUST - EQAS REPORT

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Melakidaram	Details About Robust Analysis
Lab No	9260	Detail About Monthly Summary
Month	August	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 23/08/2023


Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Robust Accuracy	Warning Signal	SDI	Rank
1	GLUCOSE III	GOD-POD III	Biosystems	34	172.03	5.51	9.48	176.9 mg/dL	0.51	3.
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	1.76	6.47	0.11	172 mg/dL	-0.35	0.
3	CHOLESTROL III	CHOD-PAP III	Biosystems	33	114.98	3.90	4.49	115.5 mg/dL	0.12	1.

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results
Yearly Summary	Change Password	Home


மருத்துவ அலுவலர்
அரசு ஆரம்ப சுகாதார நிலையம்
மேலக்கிடாரம்
இராமநாதபுரம் மாவட்டம்