

2023. JULY EQAS REPORT

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Melakidaram
 Lab No 9260
 Month July
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis

[Detail About Monthly Summary](#)

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 17/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	108.24	4.71	5.10	104.5 mg/dL	-0.73	1.70
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	5.84	5.91	0.34	6.42 mg/dL	1.68	0.12
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	113.73	3.56	4.04	114.9 mg/dL	0.29	1.39

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

11.08.23
 மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்
 சென்னை-16
 இலாபநிழலற்ற மருந்துகள்