

2023 - APRIL EQAS REPORT



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name	Govt. PHC, Melakidaram	Lab No	9260
Constituent Group	Chemistry III	Date of Result Entered :	12/04/2023
PT item	Lyophilized human serum based	Date of Report Published :	06/05/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	35	118.67	7.30	8.67	116.6 mg/dL	-0.24	2.93
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	34	2.18	6.68	0.15	2.26 mg/dL	0.55	0.05
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	112.74	5.35	6.03	107.4 mg/dL	-0.89	2.07

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Melakidaram

GOVT. MEDICAL OFFICER, GOVT. PRIMARY HEALTH CENTRE, MELAKIDARAM, KADALADI TALUK, RAMANATHAPURAM

Melakidaram

MILNADU623528

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

(Signature)
மருத்துவ அலுவலர்
சுக அலுவலர் கல்யாண குமாரன்
மெலக்கிடாரம்
மீலநாடு 623528