



G MULTI SERVICES

78, Yashob Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin: 638 452.
 gmultiservicesgobi@gmail.com. Cell : 98897 20654.
 Bio Medical Instruments Services

Dr. A. Anitha Sanyal MBBS, BSc (CC) (ICE)
 Bio Medical Engineer

Calibration Protocol & Report for Medical Equipment

Equipment : SEMI AUTO ANALYSER PROTECT CLASS /TYPE : Risk Class I
 MAKE : ROBOYIK S. No. : ATCG214322GRBK
 MODEL : PRIEST TOUCH Location : UPHC, THIRUPATHUR

S.No	Measurement Parameters		Status	Remarks
1	Electrical Test	(a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	230 volts 0 to 5 volts No Zero ohms 10mA NA	Stabilizer/UPS Well in Ranges Well
2	Physical Condition Test	(a) Cabinet & Visible damages, if any (b) Front Panel/Top lid, buckets, motor assembly & power cord (c) Risk of mechanical damages, if any	No OK No	Checked OK
3	Performance Test	(a) On /Off switch (b) switch & functional tests (c) Running test (d) Abnormalities (e) Display checkup (f) Keypad/indicators & connector	Normal OK OK No Well Well	OK Perfect Perfect
4	Calibration Test	(a) End Point Test (b) Display checkup (c) Filter selection (d) Record OD (e) Calculation (f) Two Point Test (g) Kinetic Test (h) Result	Done OK OK OK Well Perfect	Perfect
Formulated By: Edition Research Laboratory		Issued By:	User Department:	
Endorsed By : Research team head				
Reviewed On: 04.09.23 Next Review : 03.09.24				

Karunika
 MEDICAL OFFICER
 Urban Primary Health Centre,
 Tirupattur - 635 601.