



G MULTI SERVICES

7B, Yashob Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin: 638 452.
 gmulti@servicesgobichettipalayam.com Cell: 98897 20654.
 Bio Medical Instruments Services

Er. A. Nathul Rasiq DECE., BE (ICE)
 Bio Medical Engineer

Calibration Protocol & Report for Medical Equipment

Equipment : CENTRIFUGE
 MAKE : REMI
 MODEL : C-852

PROTECT CLASS /TYPE : Risk Class I
 S. No. : ZCLN-37011
 Location : UPHC, THIRUPATHUR

S.No	Measurement Parameters		Status	Remarks
1	Electrical Test	(a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/Tire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	230 volts 0 to 5 volts No Zero ohms 10mA NA	Stabilizer/UPS Within Ranges Well
2	Physical Condition Test	(a) Cabinet & Visible damages, if any (b) Front Panel/Top lid, buckets, motor assembly & power cord (c) Risk of mechanical damages, if any	No OK No	Checked OK
3	Performance Test	(a) On /Off switch (b) Speed regulator switch & functional tests (c) Running test (d) Abnormalities (e) Display checkup (f) Keypad/indicators & connector	Normal OK Accurate OK No Well Well	Ok Perfect Perfect
4	Calibration Test	(a) Vibration test (b) Speed regulation gradual hike/wobbling test (c) Running test with load (d) Display checkup (e) Timer test	Done Ok	Perfect
Formulated By Edition Research Laboratory		Issued By:	User Department:	
Endorsed By : Research team head				
Reviewed On: 04.09.23 Next Review : 03.09.24				

Kanulian
 MEDICAL OFFICER
 Urban Primary Health Centre
 Tirupattur - 635 601.