

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Latheri
 Lab No 13041
 Month March
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No. of Participants | DV | Participants | | Your Value | SDI | Q |
|-------|----------------|---------------------------------------|--|---------------------|--------|--------------|------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1878 | 91.48 | 7.46 | 6.82 | 91 mg/dL | -0.07 | 0.31 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1851 | 1.33 | 9.08 | 0.12 | 1.4 mg/dL | 0.58 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1761 | 110.93 | 5.23 | 5.80 | 115 mg/dL | 0.70 | 0.28 |

| SDI Range | Interpretation |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ±1.01 to ±2.00 | Good. |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. |
| Beyond ±3.0 | Unacceptable performance. Action Signal. |

Links

| | | | |
|--------------------------------|---------------------------------|------------------------------------|--|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Sign-out |