



RAJA MEDICAL SYSTEMS

No.15, V.O.C STREET, COIMBATORE - 641 011. MOBILE : 96263 1379

Calibration Protocol & Report for Medical Equipment

Equipment : SEMI AUTO ANALYSER

MAKE : ROBONIK

MODEL : PRIETEST TOUCH

PROTECT CLASS /TYPE : Risk Class I

S. No. : AT2800418R8K

Location : GPHC, NEDUGULA

S.No	Measurement Parameters	Status	Remarks
1	Electrical Test (a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	230 volts 0 to 5 volts No Zero ohms 10mA NA	Stabilizer/UPS With in Ranges Well
2	Physical Condition Test (a) Cabinet & Visible damages, if any (b) Front Panet/Top lid, buckets, motor assembly & power cord (c) Risk of mechanical damages, if any	No OK No	 Checked OK
3	Performance Test (a) On /Off switch (b) switch & functional tests (c) Running test (d) Abnormalities (e) Display checkup (f) Keypad/indicators & connector	Normal OK OK No Well Well	Ok Perfect Perfect
4	Calibration Test (a) End Point Test (b) Display checkup (c) Filter selection (d) Record OD (e) Calculation (f) Two Point Test (g) Kinetic Test (h) Result	Done Ok OK OK Well Perfect	Perfect
Formulated By: Einstein Research Laboratory		Issued By:	User Department:
Endorsed By : Research team head		RAJA MEDICAL SYSTEMS No.15, V.O.C STREET, COIMBATORE - 641011.	
Reviewed On: 10.02.23 Next Review : 09.02.24			