



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 2399

Distribution No.: 160-E

Month/Year: May/2023

Instrument ID: Mindray BC-5130 (TR-0C005114)

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 26-07-2023[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	3.19	3.13	6.32	7.35	0.070	-0.57	0.06	0.1	0.006	-0.42
RBC x10 ⁶ /µl	1	4.12	4.09	8.21	8.09	0.009	0.54	0.03	0.04	0.002	-0.27
Hb g/dl	1	11.6	11.5	23.1	22.9	0.027	0.27	0.1	0.1	0.007	0.00
HCT%	1	37.1	36.8	73.9	73.8	0.184	0.02	0.3	0.4	0.024	-0.34
MCV-fl	1	90.1	90.1	180.2	182.45	0.353	-0.22	0	0.3	0.022	-0.81
MCH-Pg	1	28.1	28.1	56.2	56.4	0.075	-0.11	0	0.2	0.013	-0.90
MCHC-g/dl	1	31.2	31.2	62.4	61.45	0.149	0.24	0	0.3	0.016	-1.01
Plt. x10 ³ /µl	1	260	251	511	422	1.693	1.88	9	6	0.326	0.58
Retic %	2										

P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
DLC%	3	Nrbcs=1 , Poly=01 L=02, E=01, Mono/Promono=01 , B1=95 P.M.=00, Mye=00, Meta=00, Other=
RBC Morphology	3	Blast: 75-94, Lympho: 4-12, Poly: 2-5, nRBC/ Mono/Eos/Baso/Myelo/Meta/ Promyelo: 0-5
Diagnosis	3	Predominantly: Normocytic/ Normochromic, Moderate: Anisocytosis, Microcytic, poikilocytosis
		Acute leukemia.
		Acute Leukemia (AL)

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 160--E	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10³/µl	1	314	313	90.1	92.33	4.47	1.6	5.43	6.07
RBC x10⁶/µl	1	314	314	85.99	90.76	4.46	3.18	9.55	6.06
Hb g/dl	1	314	314	87.9	92.04	4.46	2.87	7.64	5.09
HCT%	1	314	313	92.01	91.37	4.15	3.51	3.84	5.12
MCV-fl	1	314	312	94.87	93.59	3.21	1.28	1.92	5.13
MCH-Pg	1	314	312	88.46	92.31	4.81	2.88	6.73	4.81
MCHC-g/dl	1	314	312	91.35	87.82	4.49	4.49	4.16	7.69
Plt. x10³/µl	1	314	313	94.89	92.97	3.19	4.15	1.92	2.88
ReticCount%	2	314	259	91.12	81.85	5.02	11.2	3.86	6.95
PS Assessment	3	314	281	Satisfactory :97.14%, Borderline Sat. :1.91%, Unsatisfactory :0.95%					

***Comments:**

1). Among Lab (EQA) : Results acceptable.

2). Within Lab (IQA) : Precision acceptable.

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ± 2 : Acceptable, Z score ± 2 to ± 3 :Warning Signal, Z score $> \pm 3$: Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between " 0 to ± 2 " are texted in green colour. Z score value between " ± 2 to ± 3 " are texted in orange colour. Z score value $> \pm 3$ are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value ($0.3*SDPA$). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value ($0.3*SDPA$).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----