

Calibration Certificate Biochemistry Analyzer

Name	Biochemistry Analyzer			
Hospital	PRIMARY HEALTH CENTRE (EZHICHUR).			
Date of Installation	25/09/2023			
Instrument identification				QC Passed
Serial no.	ATCD0400519RBK	Date of Calibration	25/09/2023	
Model	Prietest Touch	Calibration Due Date	23/09/2024	

Inspection LIST (Relevant information should be filled in)

	YES/NO
1. The Voltage of Instrument should be 220V+10%	Yes
2. Ground Voltage should be less than 10 volts.	Yes
3. The instrument is tested with the simulator with all leads connected.	-
4. AC frequency should be 50 Hz and all filters are in ON condition.	Yes
5. QC data and live reaction curve working condition is Ok	Yes
6. All values working good	Yes

Quality Control Result	Result	Remark
Checked instrument with Quality control, working Satisfactorily	Ok	

Test results is coming ok	✓	Remarks	Done
---------------------------	---	---------	------

Signature of Technical person

