



Vanavil Calibrations
Make It Simple, But Significant !

CALIBRATION CERTIFICATE

Page no: 1 of 1

Customer Name & Address: M/S GOVERNMENT PRIMARY HEALTH CENTRE, SERAPATTU, KALLAKURICHI - 606 401.		Certificate No: VCPL/CC/23-24/31-003	
		Date of Calibration: 27-09-2023	
		Date of issue: 28-09-2023	
		Date of Receipt: 27-09-2023	
		Next cal Date: 26-09-2024	
Details of Device Under Calibration			
Description:	MICROPIPETTE	Make:	THERMO SCIENTIFIC
Range:	(100 to 1000) μ l	Model/Type:	FINNPIPETTE F3
Resolution:	0.2 μ l	Condition of DUC	GOOD
Serial Number	-	Location:	LABORATORY
ID number	GPHC/LAB/MP-01	Cal. At:	LAB
Environmental Conditions		Calibration Method Used	
Temperature	23.5 °C	Calibration Method Used	ISO 8655:2002 PART-6
Humidity	51 % Rh	Cal procedure no:	-

Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTECH	15-03-2024

RESULT OF CALIBRATION

S.NO	Duc Readings (μ l)	Standard Reading Mean Value (mg)	ERROR (mg)
1	100.0	99.89	0.11
2	500.0	499.84	0.16
3	1000.0	999.79	0.21

Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

C. Akash

C.AKASH

(CALIBRATION ENGINEER)



End of the Report ***

Authorised by:

S. Lakshmanan

S. LAKSHMANAN MURUGAN,
Technical Manager
Vanavil Calibrations Pvt Ltd.

Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

52, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2nd Floor, KVR Flats, 2nd Cross Street, Thangam Nagar, Pallikaranai, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com