



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

All Analyser Summary Report - May 2023

Lab Name	Govt. PHC, Pongampattu	Lab No	13030
Constituent Group	Chemistry III	Date of Result Entered	17/05/2023
PT item	Lyophilized human serum based	Date of Report Published	05/06/2023

Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	33	179.81	5.48	9.85			3.43
			Transasia / Erba	19	177.53	5.62	9.98			4.58
			Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	175.35 mg/dL	-0.59	0.44
			Any Analyser (Automation / Semi Automation)	69	181.20	5.81	10.52			2.53
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.78	8.91	0.16			0.05
			Transasia / Erba	22	1.85	7.73	0.14			0.06
			Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.79 mg/dL	0.23	0.01
			Any Analyser (Automation / Semi Automation)	68	1.75	7.05	0.12			0.03
3	CHOLESTROL III	CHOD-PAP III	Biosystems	36	116.50	6.42	7.47			2.49
			Transasia / Erba	20	115.73	3.06	3.54			1.58
			Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	114.51 mg/dL	-0.02	0.27
			Any Analyser (Automation / Semi Automation)	154	114.84	5.08	5.83			0.94

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.

Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS:

Govt. PHC, Pongampattu

Pongampattu

TAMILNADU,

Pamela Christudoss

Coordinator Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****