

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



12132

Lab Name Govt. PHC, Pikkili Lab No

Constituent Group Chemistry III Date of Result Entered : 20/06/2023

PT item Lyophilized human serum based Date of Report Published : 04/07/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	SDI	11
						CV	SD	Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	298.5 mg/dL	0.80	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.75 mg/dL	1.05	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	113.1 mg/dL	-0.12	0.24

SDI Range	Interpretation			
Within -1.00 to +1.00	Excellent.			
Within ±1.01 to ±2.00	Good.			
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.			
Beyond ±3.0	Unacceptable performance. Action Signal.			

LAB ADDRESS:

Govt. PHC, Pikkili

THE MEDICAL OFFICER, PIKKILI PHC, PENNAGARAM, DHARMAPURI DIST.

Pikkili

TAMILNADU636808

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102 Panela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator

Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report *******