

## **CHRISTIAN MEDICAL COLLEGE**

## **DEPARTMENT OF CLINICAL BIOCHEMISTRY**

## **CMC EXTERNAL QUALITY ASSURANCE SCHEME**





Lab Name Govt. PHC, Pikkili Lab No 12132

Constituent Group Chemistry III Date of Result Entered : 20/03/2023

PT item Lyophilized human serum based Date of Report Published : 04/04/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	SDI	
SI.NO						CV	SD	Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	86.2 mg/dL	-0.77	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.29 mg/dL	-0.33	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	108 mg/dL	-0.50	0.28

SDI Range	Interpretation			
Within -1.00 to +1.00	Excellent.			
Within ±1.01 to ±2.00	Good.			
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.			
Beyond ±3.0	Unacceptable performance. Action Signal.			

LAB ADDRESS:

Govt. PHC, Pikkili

THE MEDICAL OFFICER, PIKKILI PHC, PENNAGARAM, DHARMAPURI DIST.

Pikkili

TAMILNADU636808

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102 Panela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator

**Christian Medical College, Vellore** 

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

\*\*\*\*\*\*\*\*\* End of Report \*\*\*\*\*\*\*