



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name: Govt. PHC, Vijayamangalam

Lab No: 12200

Constituent Group: Chemistry III

Date of Result Entered: 20/03/2023

PT item: Lyophilized human serum based

Date of Report Published: 04/04/2023

S.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	92.67	8.46	7.84	98 mg/dL	0.68	2.61
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.37	9.41	0.13	1.1 mg/dL	-2.09	0.04
3	CHOLESTROL III	CHOD-PAP III	Biosystems	37	112.13	5.96	6.68	104 mg/dL	-1.22	2.20
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

2 score -> 1.2

### LAB ADDRESS :

Govt. PHC, Vijayamangalam  
THE MEDICAL OFFICER, PERUNTHURAI  
Vijayamangalam  
TAMILNADU

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Handwritten signature*  
மரு. அன்பு சந்திரா பிஷையா  
விவரம் - 638056

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*