



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name: Govt. PHC, Vijayamangalam
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No: 12200
 Date of Result Entered: 20/05/2023
 Date of Report Published: 05/06/2023

Sl.No	Analyte	Method/ Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	33	179.81	5.48	9.85	195 mg/dL	1.54	3.43
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.78	8.91	0.16	1.6 mg/dL	-1.13	0.05
3	CHOLESTROL III	CHOD-PAP III	Biosystems	36	116.50	6.42	7.47	110 mg/dL	-0.87	2.49
SDI Range										
				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

2 score → 1.18

LAB ADDRESS :
 Govt. PHC, Vijayamangalam
 THE MEDICAL OFFICER, PERUNTHURAI
 Vijayamangalam
 TAMILNADU

Coordinator Contact Details:
 Email: clingdc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Handwritten signature
 சி.சி.சி. மருத்துவ கல்லூரி
 மருத்துவ அலுவலகம் - 638056,
 விஜயமங்கலம்

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****