



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name

Govt. PHC, Vijayamangalam

Lab No

12200

Constituent Group

Chemistry III

Date of Result Entered :

20/06/2023

PT item

Lyophilized human serum based

Date of Report Published :

04/07/2023

SL No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	280.51	4.07	11.42	275 mg/dL	-0.48	3.81
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	39	2.67	6.50	0.17	2.4 mg/dL	-1.55	0.06
3	CHOLESTROL III	CHOD-PAP III	Biosystems	37	113.79	4.23	4.81	112 mg/dL	-0.37	1.58
SDI Range										
				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

Z Score → 0.8

LAB ADDRESS :

Govt. PHC, Vijayamangalam
THE MEDICAL OFFICER, PERUNTHURAI
Vijayamangalam
TAMILNADU

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christodoss
20.06.2023
அரசு ஆரம்ப சுகாதார நிலையம்
விஜயமங்கலம் - 638056.

Dr. Pamela Christodoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****