



**CHRISTIAN MEDICAL COLLEGE**  
DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JULY 2023



Lab Name Govt. PHC, Vijayamangalam  
Constituent Group Chemistry III  
PT item Lyophilized human serum based

Lab No 12200  
Date of Result Entered : 20/07/2023  
Date of Report Published : 07/08/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	108.24	4.71	5.40	105 mg/dL	-0.64	1.70
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	5.84	5.91	0.34	5.6 mg/dL	-0.70	0.12
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	113.73	3.56	4.04	115 mg/dL	0.31	1.39

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

LAB ADDRESS :  
Govt. PHC, Vijayamangalam  
THE MEDICAL OFFICER, PERUNTHURAI  
Vijayamangalam  
TAMILNADU

Coordinator Contact Details:  
Email: clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

Pamela Christudoss  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

மேலே உள்ள விவரம்  
அரசு ஆரம்ப சுகாதார நிலையம்  
விஜயமங்கலம் - 638056.  
Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not subcontract any components  
\*\*\*\*\* End of Report \*\*\*\*\*