



Lab Name: Govt. PHC, Vijayamangalam

Constituent Group: Chemistry III

PT item: Lyophilized human serum based

Lab No: 12200

Date of Result Entered: 20/04/2023

Date of Report Published: 06/05/2023

CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	35	118.67	7.30	8.67	127 mg/dL	0.96	2.93
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	34	2.18	6.68	0.15	1.9 mg/dL	-1.92	0.05
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	112.74	5.35	6.03	107 mg/dL	-0.95	2.07
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

2 SCORE → 1.2

LAB ADDRESS :
Govt. PHC, Vijayamangalam
THE MEDICAL OFFICER, PERUNTHURAI
Vijayamangalam
TAMILNADU

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Palanivel
மாணவியல் அலுவலர்
மாணவியல் அலுவலர் - 638056

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****