

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY CMC EXTERNAL QUALITY ASSURANCE SCHEME MONTHLY SUMMARY REPORT - AUGUST 2023



Govt. PHC, Vijayamangalam

Lyophilized human serum based

12200 21/08/2023

Constituent Group

PT item

Chemistry III

Date of Result Entered : Date of Report Published :

01/09/2023

51.No	Analyte	Method / Principle Name	Analyzer Name	No of	DV	Participants Participants		Your Value	SDI	
e in the				Participants		CV	SD)		SIL	34.8
1	GLUCOSE III	GOD-POD III	Biosystems	34	172.03	5.51	9.48	175 mg/dL	0.31	3.25
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	1.76	6.47	0.11	2 mg/dL	2.11	0.04
3	CHOLESTROL III	CHOD-PAP III	Biosystems	33	114.98	3.90	4.49	109 mg/dL	-1.33	1.56

SDI Range	Interpretation				
Within -1.00 to +1.00	Excellent.				
Within ±1.01 to ±2.00	Good.				
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.				
Beyond ±3.0	Unacceptable performance. Action Signal.				

LAB ADDRESS: Govt. PHC, Vijayamangalam THE MEDICAL OFFICER, PERUNTHURAI Vijayamangalam TAMILNADU

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components

Panola Christudoss

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore