



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MAY 2023



Lab Name Govt. Block PHC, Nedugula Lab No 7541  
Constituent Group Chemistry III Date of Result Entered : 21/05/2023  
PT item Lyophilized human serum based Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	169 mg/dL	-1.26	0.44
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	163	27.91	13.39	3.74	22.7 mg/dL	-1.39	0.59
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.9 mg/dL	1.06	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	119 mg/dL	0.76	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :  
Govt. Block PHC, Nedugula  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, NILGIRIS DISTRICT.  
Nedugula  
TAMILNADU643217

*[Signature]*  
MEDICAL OFFICER  
Govt. Primary Health Centre,  
Nedugula - 643 217.

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
- Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*