



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MARCH 2023



Lab Name Govt. Block PHC, Nedugula

Lab No 7541

Constituent Group Chemistry III

Date of Result Entered : 20/03/2023

PT item Lyophilized human serum based

Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	97 mg/dL	0.81	0.31
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	175	55.20	11.70	6.46	61 mg/dL	0.90	0.98
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.39 mg/dL	0.50	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	114 mg/dL	0.53	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :  
 Govt. Block PHC, Nedugula  
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, NILGIRIS DISTRICT.  
 Nedugula  
 TAMILNADU643217

Coordinator Contact Details:  
 Email:clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*S. J. S.*  
 MEDICAL OFFICER  
 Govt. Primary Health Centre,  
 Nedugula - 643 217

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*