

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Poygai
 Lab No: 10652
 Month: April
 Year: 2023
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	124 mg/dL	0.63	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.3 mg/dL	0.80	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	105 mg/dL	-1.14	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links	Modify Address	Modify Methods	Submit Lab Results
Yearly Summary	Change Password	Home	View Lab Monthly Summary

Signature
 மருத்துவ அலுவலர்
 சி. சாதார நிலைய
 அரசு ஆரோக்கிய அமைச்சு
 மதுரை - 625 002