

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Poygai
 Lab No: 10652
 Month: August
 Year: 2023
 Constituent: Chemistry III
 Group:
 Click on the analyte to view Graphical Data

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

All Analyser Result

Date of Result Entered : 20/08/2023

Date of Report Published : 01/09/2023

Analyte	Method Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
1 GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	175 mg/dL	-0.06	0.42
2 CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.75 mg/dL	-0.08	0.01
3 CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	114 mg/dL	0.04	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Handwritten signature and stamp at the bottom right corner.