

03/12/2023

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Poygai
 Lab No 10652
 Month June
 Year 2023
 Constituent Chemistry III
 Group

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl No	Analyte	Method Principle Name	Analyzer Name	No of Participants	CV	CV	CV	CV	CV	CV
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	314 mg/dL	1.85	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.52 mg/dL	-0.29	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	112 mg/dL	-0.34	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்
 - 636 116