

8-915519
9.0565875

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Poygai
Lab No: 10652
Lab No: March
Month: 2023
Year: Chemistry III
Client: Click on the analyte to view Graphical Data
Group:

Details About Robust Analysis
Detail About Monthly Summary
Detail about SDI

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

Sl. No	Participant Name	Analyte Name	No of Participants	GV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	97 mg/dL	0.81	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.32 mg/dL	-0.08	0.01
3	CHOLESTROL III	CHOD-PAP III Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	108 mg/dL	-0.50	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Handwritten Signature
மருத்துவ அலுவலர்
பிரதீப கண்ணன்
முனைவர் - 2023