





CALIBRATION /TESTING REPORT					
CUSTOMER DETAILS: GOVT URBAN PRIMARY HEALTH CENTRE ARUPPUKOTTAI TAMIL NADU- 626 101		CERTIFICATE NO	TICS/UR-APK/23-24/012		
		EQUIPMENT NAME	CENTRIFUGE		
		MANUFACTURE	BIONIX		
		MODEL	BIO CHEMISTRY ANALYZER		
CALIBRATION DATE	13.07.2023	SERIAL NUMBER	BMS/10/02		
CALIBRATION DUE DATE	12.07.2024	INSTALLED DEPARTMENT	LAB		
<p>1. General maintenance required <input type="checkbox"/> Yes <input checked="" type="checkbox"/> NO</p> <p>2. Any Physical damage observed <input type="checkbox"/> Yes <input checked="" type="checkbox"/> NO</p> <p>3. Spare Replacement (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> NO</p>					
TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER		
DEGTAL TECHOMETER	KUSAM MECO	KM2234 BL	S1049875		
ELECTRICAL SAFTY ANALIZER	DATREND	V Pad mini	VPM 18040011		
TEST REPORT - PERFORMANCE					Temperature 25°C Humidity RH55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERNESS	STATUS
ROTATION PER MINUTE	RPM	SPEED 2	1650	+5	PASS
		SPEED 5	2650	±5	PASS
<small>*The above mentioned instrument has been tested using standards manufacturers recommended protocols, equipments having traceability to National/ International standards.</small>					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE		<input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED		<input type="checkbox"/> REMOVED FROM USE	
Calibrated By			Authorized Sign With Seal		
					

*Electrical safety test result attached