





CALIBRATION /TESTING REPORT					
CUSTOMER DETAILS:		CERTIFICATE NO	TICS/UR-APK/23-24/015		
GOVT URBAN PRIMARY HEALTH CENTRE		EQUIPMENT NAME	MICRO PIPETTE		
ARUPPUKOTTAI		MANUFACTURE	TERMO		
TAMIL NADU- 626 101		MODEL	BIO CHEMISTRY ANALYZER		
CALIBRATION DATE	13.07.2023	SERIAL NUMBER	RW02994		
CALIBRATION DUE DATE	12.07.2024	INSTALLED DEPARTMENT	LAB		
<p>1. General maintenance required <input type="checkbox"/> Yes <input checked="" type="checkbox"/> NO</p> <p>2. Any Physical damage observed <input type="checkbox"/> Yes <input checked="" type="checkbox"/> NO</p> <p>3. Spare Replacement (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> NO</p>					
TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER		
ELECTRICAL SEMI MICRO BALANCE	SHIMADZU	TVCSPL22/07/884-01	27072023		
TEST REPORT - PERFORMANCE					Temperature 25°C Humidity RH55%
PARAMETERS	UNITS	SET VALUES	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	L	1	10.2	±1	PASS
		49	49.3	±1	PASS
<small>*The above mentioned instrument has been tested using standards manufacturers recommended protocols, equipments having traceability to National/ International standards.</small>					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By			Authorized Sign With Seal		
					

*Electrical safety test result attached