

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. Block PHC, Gandamanur	<a href="#">Details About Robust Analysis</a>									
Lab No	7218	<a href="#">Detail About Monthly Summary</a>									
Month	August	<a href="#">Detail about SDI</a>									
Year	2023										
Constituent Group	Chemistry III										
		<a href="#">Click on the analyte to view Graphical Data</a>				<a href="#">All Analyser Result</a>		<a href="#">Print</a>			
		Date of Result Entered : 19/08/2023									
		Date of Report Published : 01/09/2023									
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U	
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	164 mg/dL	-1.29	0.42	
2	<a href="#">UREA III</a>	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	174	28.29	15.20	4.30	34.8 mg/dL	1.51	0.65	
3	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.6 mg/dL	-1.34	0.01	
4	<a href="#">URIC ACID III</a>	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	40	5.43	10.18	0.55	5.3 mg/dL	-0.24	0.17	
5	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	106 mg/dL	-1.68	0.22	
6	<a href="#">TRIGLYCERIDE III</a>	GPO-PAP / Enzymatic Colorimetric / End Point III	Robonik ( Semi automatic / Fully Automatic )	11	137.31	8.60	11.80	124 mg/dL	-1.13	7.12	
SDI Range		Interpretation									
Within -1.00 to +1.00		Excellent.									
Within ±1.01 to ±2.00		Good.									
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.									
Beyond ±3.0		Unacceptable performance. Action Signal.									

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>