



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name Govt. PHC, Chinnna Mottur

Lab No

13038

Constituent Group Chemistry III

Date of Result Entered :

19/08/2023

PT Item Lyophilized human serum based

Date of Report Published :

01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	185 mg/dL	1.05	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.6 mg/dL	-1.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	100 mg/dL	-2.98	0.22
<b>SDI Range Interpretation</b>										
Within -1.00 to +1.00				Excellent.						
Within $\pm 1.01$ to $\pm 2.00$				Good.						
Within $\pm 2.01$ to $\pm 2.99$				Accept with caution. Warning Signal.						
Beyond $\pm 3.0$				Unacceptable performance. Action Signal.						

## LAB ADDRESS :

Govt. PHC, Chinnna Mottur

Chinnna Mottur

TAMILNADU

Coordinator Contact Details:  
Email: cllngc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential