

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt PHC, Kesankadu
 Lab No: 12550
 Month: July
 Year: 2023

Chemistry Group: Chemistry III

Click on the analyte to view Graphical Data

- Details About Robust Analysis
- Detail About Monthly Summary
- Detail about SDI

Date of Result Entered : 15/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analysar Name	No of Participants	DV	CV	SD	Year Value	SDI	Z
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	117.3 mg/dL	1.31	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Pictate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	8.71 mg/dL	-0.34	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.86	114.2 mg/dL	0.26	0.22

SDI Range		Interpretation
Within -1.00 to +1.00		Excellent.
Within ±1.01 to ±2.00		Good.
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.
Beyond ±3.0		Unacceptable performance. Action Signal.